Track Light Template

Forest
DENTALEZ

ALERT SYMBOL KEY The safety notices associated with the following should be given special attention when they appear in maintenance, operating and emergency procedures in the guide.

⚠ CAUTION indicates that a particular procedure or precaution must be followed to avoid system irregularity or possible damage to the product.

⚠ WARNING indicates that the personal safety of the patient, end user or technician could be compromised by disregarding the WARNING. Not following instructions may result in an injury.

⚠ *WARNING: We recommend consulting a licensed professional, i.e. Structural or Architectural Engineer when attaching a fixture to the building structure.*

⚠ *CAUTION: Recommended and provided mounting hardware may not be suitable for your specific installation due to ceiling type. Please use caution when selecting hardware due to supporting structure type to avoid possible damage. Forest is not liable for damage due to hardware choice.*

⚠ *CAUTION: Emailed, faxed or copied templates may jeopardize precise full scale measurements.*

Recommended Hardware: 1/4" x 3" Hex Head Lag Screw, Zinc, 8 each; Spacer, 8 each.



DTE OREGON, INC. *DENTALEZ.COM*
6200 NE Cherry Drive, Hillsboro, OR 97124 USA (TF) 800 . 423 . 3555 (T) 503 . 640 . 3012 (F) 503 . 693 . 9715